



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

### **Bravo Technical Services, Inc.**

**130 Johnson Drive  
Terre Haute, IN 47802**

Fulfills the requirements of

### **ISO/IEC 17025:2017**

and national standard

### **ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

### **CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 04 January 2024  
Certificate Number: AC-1300



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)**

**Bravo Technical Services, Inc.**

130 W. Johnson Drive

Terre Haute, IN 47802

Richard McClain 812-235-3757

bmcclain@bravots.com www.bravots.com

**CALIBRATION**

Valid to: **January 4, 2024**

Certificate Number: **AC-1300**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (333 to 1 000) V	18 $\mu$ V 0.2 mV 2.3 mV 24 mV 54 mV	Fluke 5522A Multiproduct Calibrator
DC Current – Source	Up to 330 $\mu$ A (0.33 to 3.3) mA (3.3 to 33) mA 33 mA to 1.1 A (1.1 to 3) A	0.22 $\mu$ A 2.1 $\mu$ A 25 $\mu$ A 0.4 mA 1.3 mA	Fluke 5522A Multiproduct Calibrator
AC Voltage – Source	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.13 V 68 $\mu$ V 68 $\mu$ V 81 $\mu$ V 0.22 mV 1 mV 1.3 mV 0.55 mV 0.55 mV 0.59 mV 1.5 mV 9.3 mV	Fluke 5522A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	(0.33 to 3.3) V		Fluke 5522A Multiproduct Calibrator
	(10 to 45) Hz	13 mV	
	45 Hz to 10 kHz	7.8 mV	
	(10 to 20) kHz	7.9 mV	
	(20 to 50) kHz	8.5 mV	
	(50 to 100) kHz	17 mV	
	(100 to 500) kHz	0.19 V	
	(3.3 to 33) V		
	(10 to 45) Hz	0.13 V	
	45 Hz to 10 kHz	66 mV	
	(10 to 20) kHz	79 mV	
	(20 to 50) kHz	82 mV	
	(50 to 100) kHz	0.15 V	
	(33 to 330) V		
	10 Hz to 1 kHz	0.49 V	
	(1 to 10) kHz	0.65 V	
	(10 to 20) kHz	0.81 V	
	(20 to 50) kHz	0.95 V	
	(50 to 100) kHz	2 V	
	(330 to 1 020) V		
(10 to 45) Hz	2.5 V		
45 Hz to 1 kHz	1.1 V		
(1 to 5) kHz	1.3 V		
(5 to 8) kHz	1.4 V		
AC Current – Source	(33 to 330) $\mu$ A		Fluke 5522A Multiproduct Calibrator
	(10 to 20) Hz	1.9 $\mu$ A	
	(20 to 45) Hz	1.7 $\mu$ A	
	45 Hz to 1 kHz	1.1 $\mu$ A	
	(1 to 5) kHz	2.9 $\mu$ A	
	(5 to 10) kHz	4.2 $\mu$ A	
	(10 to 30) kHz	7.1 $\mu$ A	
	(0.33 to 3.3) mA		
	(10 to 20) Hz	20 $\mu$ A	
	(20 to 45) Hz	16 $\mu$ A	
	45 Hz to 1 kHz	8.7 $\mu$ A	
	(1 to 5) kHz	24 $\mu$ A	
	(5 to 10) kHz	26 $\mu$ A	
	(10 to 30) kHz	33 $\mu$ A	



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	(3.3 to 33) mA		Fluke 5522A Multiproduct Calibrator
	(10 to 20) Hz	0.19 mA	
	(20 to 45) Hz	0.16 mA	
	45 Hz to 1 kHz	75 $\mu$ A	
	(1 to 5) kHz	41 mA	
	(5 to 10) kHz	0.28 mA	
	(10 to 30) kHz	0.48 mA	
	(33 to 330) mA		
	(10 to 20) Hz	1.9 mA	
	(20 to 45) Hz	2 mA	
	45 Hz to 1 kHz	1.6 mA	
	(1 to 5) kHz	13 mA	
	(5 to 10) kHz	13 mA	
	(10 to 30) kHz	14 mA	
	(0.33 to 3) A		
(10 to 45) Hz	5.4 mA		
45 Hz to 1 kHz	4.3 mA		
(1 to 5) kHz	16 mA		
(5 to 10) kHz	40 mA		
Resistance – Source	Up to 11 $\Omega$	13 m $\Omega$	Fluke 5522A Multiproduct Calibrator
	(11 to 33) $\Omega$	20 m $\Omega$	
	(33 to 110) $\Omega$	23 m $\Omega$	
	(110 to 330) $\Omega$	44 m $\Omega$	
	(0.33 to 1.1) k $\Omega$	94 m $\Omega$	
	(1.1 to 3.3) k $\Omega$	0.44 $\Omega$	
	(3.3 to 11) k $\Omega$	0.87 $\Omega$	
	(11 to 33) k $\Omega$	4.8 $\Omega$	
	(33 to 110) k $\Omega$	14 $\Omega$	
	(3.3 to 11) k $\Omega$	70 $\Omega$	
	(0.33 to 1.1) M $\Omega$	0.22 k $\Omega$	
	(1.1 to 3.3) M $\Omega$	3.4 k $\Omega$	
	(3.3 to 11) M $\Omega$	13 k $\Omega$	
	(11 to 33) M $\Omega$	0.72 M $\Omega$	
	(33 to 110) M $\Omega$	2.6 M $\Omega$	
DC Voltage – Measure	Up to 200 mV	12 $\mu$ V	Keithley 2001 7.5 Digit Multimeter
	(0.2 to 2) V	68 $\mu$ V	
	(2 to 20) V	0.7 mV	
	(20 to 200) V	10 mV	
	(200 to 1 000) V	59 mV	



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure	Up to 200 $\mu$ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A	0.13 $\mu$ A 0.97 $\mu$ A 9.6 $\mu$ A 0.12 mA 2.3 mA	Keithley 2001 7.5 Digit Multimeter
AC Voltage – Measure	Up to 200 mV (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz (0.2 to 2) V (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz (2 to 20) V (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (20 to 200) V (20 to 50) Hz 50 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (200 to 750) V (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz	0.59 mV 0.59 mV 0.15 mV 0.11 mV 0.17 mV 0.19 mV 0.73 mV 1.8 mV 5.9 mV 2.2 mV 1.5 mV 1.5 mV 1.7 mV 14 mV 7.2 mV 18 mV 59 mV 22 mV 18 mV 23 mV 31 mV 34 mV 73 mV 0.59 V 0.18 V 0.23 V 0.31 V 0.33 V 0.86 V 2.5 V 1.1 V 1.3 V	Keithley 2001 7.5 Digit Multimeter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure	Up to 200 $\mu$ A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (0.2 to 2) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (2 to 20) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (20 to 200) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (0.2 to 2) A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz	0.92 $\mu$ A 0.62 $\mu$ A 1 $\mu$ A 2.3 $\mu$ A 7.6 $\mu$ A 4.4 $\mu$ A 3.8 $\mu$ A 12 $\mu$ A 73 $\mu$ A 38 $\mu$ A 32 $\mu$ A 57 $\mu$ A 0.73 mA 0.38 mA 0.32 mA 0.67 mA 9.1 mA 5.1 mA 7.2 mA 62 mA	Keithley 2001 7.5 Digit Multimeter
DC Resistance – Measure	Up to 20 $\Omega$ (20 to 200) $\Omega$ (0.2 to 2) k $\Omega$ (2 to 20) k $\Omega$ (20 to 200) k $\Omega$ (0.2 to 2) M $\Omega$ (2 to 20) M $\Omega$ (20 to 200) M $\Omega$	18 m $\Omega$ 28 m $\Omega$ 0.28 $\Omega$ 1.7 $\Omega$ 20 $\Omega$ 0.34 k $\Omega$ 13 k $\Omega$ 2.4 M $\Omega$	Keithley 2001 7.5 Digit Multimeter

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for all parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1300.



R. Douglas Leonard Jr., VP, PILR SBU

