



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Bravo Technical Services, Inc.

**130 W. Johnson Drive
Terre Haute, IN 47802**

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President
Expiry Date: 04 January 2026
Certificate Number: AC-1300



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Bravo Technical Services, Inc.

130 W. Johnson Drive

Terre Haute, IN 47802

Richard McClain 812-235-3757

bmcclain@bravots.com www.bravots.com

CALIBRATION

Valid to: **January 4, 2026**

Certificate Number: **AC-1300**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (333 to 1 000) V	19 μ V 0.2 mV 2.3 mV 24 mV 54 mV	Fluke 5522A Multiproduct Calibrator
DC Current – Source	Up to 330 μ A (0.33 to 3.3) mA (3.3 to 33) mA 33 mA to 1.1 A (1.1 to 3) A	0.22 μ A 2.1 μ A 25 μ A 0.4 mA 1.3 mA	Fluke 5522A Multiproduct Calibrator
AC Voltage – Source	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.13 V 68 μ V 68 μ V 81 μ V 0.22 mV 1 mV 1.3 mV 0.55 mV 0.55 mV 0.59 mV 1.5 mV 9.3 mV	Fluke 5522A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	(0.33 to 3.3) V		Fluke 5522A Multiproduct Calibrator
	(10 to 45) Hz	13 mV	
	45 Hz to 10 kHz	7.8 mV	
	(10 to 20) kHz	7.9 mV	
	(20 to 50) kHz	8.5 mV	
	(50 to 100) kHz	17 mV	
	(100 to 500) kHz	0.19 V	
	(3.3 to 33) V		
	(10 to 45) Hz	0.13 V	
	45 Hz to 10 kHz	65 mV	
	(10 to 20) kHz	79 mV	
	(20 to 50) kHz	82 mV	
	(50 to 100) kHz	0.15 V	
	(33 to 330) V		
	10 Hz to 1 kHz	0.49 V	
	(1 to 10) kHz	0.65 V	
	(10 to 20) kHz	0.81 V	
	(20 to 50) kHz	0.95 V	
	(50 to 100) kHz	2 V	
	(330 to 1 020) V		
(10 to 45) Hz	2.5 V		
45 Hz to 1 kHz	1.1 V		
(1 to 5) kHz	1.3 V		
(5 to 8) kHz	1.4 V		
AC Current – Source	(33 to 330) μ A		Fluke 5522A Multiproduct Calibrator
	(10 to 20) Hz	1.9 μ A	
	(20 to 45) Hz	1.7 μ A	
	45 Hz to 1 kHz	1.1 μ A	
	(1 to 5) kHz	2.9 μ A	
	(5 to 10) kHz	4.2 μ A	
	(10 to 30) kHz	7.1 μ A	
	(0.33 to 3.3) mA		
	(10 to 20) Hz	19 μ A	
	(20 to 45) Hz	17 μ A	
	45 Hz to 1 kHz	8.7 μ A	
	(1 to 5) kHz	24 μ A	
	(5 to 10) kHz	26 μ A	
	(10 to 30) kHz	33 μ A	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	(3.3 to 33) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (33 to 330) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (0.33 to 3) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.18 mA 0.17 mA 96 μ A 41 mA 0.28 mA 0.48 mA 2.1 mA 2 mA 1.6 mA 13 mA 13 mA 14 mA 5.4 mA 4.3 mA 16 mA 40 mA	Fluke 5522A Multiproduct Calibrator
Resistance – Source	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (3.3 to 11) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω	13 m Ω 20 m Ω 23 m Ω 43 m Ω 94 m Ω 0.44 Ω 0.87 Ω 4.8 Ω 14 Ω 70 Ω 0.22 k Ω 3.4 k Ω 13 k Ω 0.72 M Ω 2.6 M Ω	Fluke 5522A Multiproduct Calibrator
DC Voltage – Measure	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 000) V	12 μ V 67 μ V 0.71 mV 10 mV 59 mV	Keithley 2001 7.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure	Up to 200 μ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A	0.13 μ A 0.97 μ A 9.6 μ A 0.12 mA 2.3 mA	Keithley 2001 7.5 Digit Multimeter
AC Voltage – Measure	Up to 200 mV (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz (0.2 to 2) V (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz (2 to 20) V (20 to 50) Hz (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (20 to 200) V (20 to 50) Hz 50 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (200 to 750) V (50 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz	0.59 mV 0.59 mV 0.15 mV 0.11 mV 0.17 mV 0.19 mV 0.73 mV 1.8 mV 5.9 mV 2.2 mV 1.5 mV 1.5 mV 1.7 mV 14 mV 7.2 mV 18 mV 59 mV 22 mV 18 mV 23 mV 31 mV 34 mV 73 mV 0.59 V 0.18 V 0.23 V 0.31 V 0.33 V 0.86 V 2.5 V 1.1 V 1.3 V	Keithley 2001 7.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure	Up to 200 μ A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (0.2 to 2) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (2 to 20) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (20 to 200) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (0.2 to 2) A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz	0.92 μ A 0.62 μ A 1 μ A 2.3 μ A 7.5 μ A 4.4 μ A 3.8 μ A 12 μ A 73 μ A 38 μ A 32 μ A 57 μ A 0.73 mA 0.38 mA 0.32 mA 0.67 mA 9.1 mA 5.1 mA 7.2 mA 62 mA	Keithley 2001 7.5 Digit Multimeter
DC Resistance – Measure	Up to 20 Ω (20 to 200) Ω (0.2 to 2) k Ω (2 to 20) k Ω (20 to 200) k Ω (0.2 to 2) M Ω (2 to 20) M Ω (20 to 200) M Ω	18 m Ω 28 m Ω 0.23 Ω 1.7 Ω 20 Ω 0.32 k Ω 12 k Ω 2.5 M Ω	Keithley 2001 7.5 Digit Multimeter

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for all parameters, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1300.



Jason Stine, Vice President

